Institut Lavoisier de Versailles

ELEMENTAL ANALYSIS (ICP-OES, AAS)



The Lavoisier Institute of Versailles is equipped with an inductively coupled plasma optical emission spectrometry apparatus (Agilent 720 Series ICP-OES).



CP-OES allows:

- a quick analysis of almost all elements of the periodic table

- the analysis of trace compounds with extremely low detection limits (of the order of g / L in dissolved phase and less than mg / kg in solid phase)

- quantitative analyzes
- complex matrix analyzes

The large dynamic range allows the ICP-OES to be able to analyze both major elements and trace elements during the same analysis without having to proceed to several dilutions. The robustness, multi-elemental character and high linearity of the ICP-OES measurement range mean that it covers a very wide range of applications; the analyzes can be carried out in very diverse types of samples (aqueous liquids or organic solvents, solutions loaded with dissolved salts, metal alloys, organic matrices, etc.) for environmental applications (soil, freshwater or water analyzes). seawater), in agribusiness (analyzes in the chemical and pharmaceutical industry, in metallurgy, but also in plants, organic tissues, wine)



Observable elements

															ICP-OES		
	Limites de Détection 720/730-ES Visée Ax														iale	•	
	Nébuliseur Ultrasonique																He
	Y																
Li	Be									•		в	С	N		F	Ne
0.05	0.003	LD : Limite de détection en µg/L à 3 x s _{blanc}															
Na	Mg	S _{blane} = écart-type du blanc sur 10 prises d'essai identiques dans des conditions AI SI P S C															
0.015	0.002	0.07 0.7 0.4 4.0															
ĸ	Са	Sc	Ti	V	Cr	Mn	Fe	Co	Ni	Cu	Zn	Ga	Ge	As	Se	Br	Kr
0.1	0.01	0.006	0.015	0.01	0.025	0.005	0.02	0.05	0.06	0.04	0.02	0.15	0.29	0.4	0.8		
Rb	Sr	Y	Zr	Nb	Мо	Тс	Ru	Rh	Pd	Ag	Cd	In	Sn	Sb	Те	1	Xe
0.4	0.002	0.01	0.01	0.04	0.04		0.08	0.07	0.16	0.04	0.012	0.28	0.15	0.3	1	-	
Cs	Ва	La	Hf	Та	W	Re	Os	Ir	Pt	Au	Hg	TI	Pb	Bi	Pa	At	Rn
	0.002	0.03	0.04	0.1	0.2	0.80	0.90	0.1	0.15	0.1	0.3	0.13	0.15	0.3			
	Rd	Ac															
				Ce	Pr	Nd	Pm	Sm	Eu	Gd	Tb	Dy	Но	Er	Tm	Yb	Lu
				0.15	0.09	0.13		0.05	0.02	0.05	•	0.03	0.04	0.06	0.03	0.003	0.004
				Th	Pa	U	Np	Pu	Am	Cm	Bk		Es	Fm	Md	Nc	Lr
				0.07		0.3											

THE BASIC ANALYSIS SERVICE OF THE LAVOISIER INSTITUTE IS OPEN TO THE ENTIRE SCIENTIFIC COMMUNITY.

Financial manager : Muriel Bouttemy

Service manager : Flavien Bourdreux